

# Declaration of Conformity



**Type of equipment:** breathalyzer  
**Brand Name /Trade Mark:** DA Tech  
**Type designation /model:** AF-35, AF-30, AF-33, ELITE, PRO X-5, FCA-35 and AM-7000  
**Manufacturer:** DA Tech Co., Ltd.

**In accordance with the following Directives:**

2014/30/EU The Electromagnetic Compatibility Directive

**The following harmonized European standards and technical specifications have been applied:**

EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements
EN 55011:2009 +A1:2010	Industrial, scientific and medical equipment. Radio-frequency disturbance characteristics. Limits and methods of measurement
EN 61326-2-2:2013	Electrical equipment for measurement, control and laboratory use. EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems
EN 61000-4-2:2009	Electromagnetic compatibility (EMC). Testing and measurement techniques. Electrostatic discharge immunity test
EN 61000-4-3:2006 +A1:2008 +A2:2010	Radiated, radio-frequency, electromagnetic field immunity test
EN 61000-4-8:2010	Electromagnetic compatibility (EMC). Testing and measurement techniques. Power frequency magnetic field immunity test

**Test report issued by:**  
EMC: CTK Co., Ltd.

**The CE Marking on the products and/or their packaging signifies that DA Tech Co., Ltd. hold the reference technical file available to the European Union authorities.**

**Place and date of issue:** 39, Pyeongcheon-ro 141 beon-fil, Bupyeong-gu, Incheon, Korea / 2017-03-09

**Authorized Signatory:**

Name:

*Kwang Hee Park*

Title:

CEO

Signature:

*K.H. Park*